



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ALARA CONSULTANTS INC.  
9524 27 Ave NW  
Edmonton, Alberta T6N1B2  
CANADA  
David Duke Phone: 780 944 2557

CALIBRATION

Valid To: January 31, 2023

Certificate Number: 4802.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,3</sup>:

I. Ionizing Radiation & Radioactivity

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Radiation Survey Instruments	1 µSv/hr to 16 mSv/hr	9.8 % of reading	JL Shepherd 28-6 calibrator
		11 % of reading	QSA Global model 773
Contamination Instruments	(1 to 9.9e <sup>6</sup> ) cpm	6.2 % of reading	Ludlum model 500-2 pulser, Ludlum model 500 pulser

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

## ALARA CONSULTANTS INC.

Edmonton, Alberta, CANADA

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 28<sup>th</sup> day of January 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 4802.01  
Valid to January 31, 2023

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*